BETTER AG
Top quality at manufacturer prices

XIQUE
FFP2/3 NR Masks
Filtering Half Mask

MODEL: LK-Z006

MODEL: LK-L005

MODEL: LK-T007

Product: FFP2 NR Mask
Plastic clips included
Stock Location EU
Individually Packed
CE certified FFP2 NR Masks

Color: White
Shelf Life: 3 years
Mask Dimensions: 15.5cm x 10.5cm

20 Mask per Box
50 Boxes per Carton
1,000 Masks per Carton

Carton Dimensions: 61x38x48cm
Carton Weight: 11kg

Product: FFP2 NR Fish Shape Mask
Plastic clips included
Stock Location EU
Individually Packed
CE certified FFP2 NR Fish Masks

Color: White
Shelf Life: 3 years
Mask Dimensions: 20.5cm x 8cm

20 Mask per Box
60 Boxes per Carton
1,200 Masks per Carton

Carton Dimensions: 53x47x53cm
Carton Weight: 12kg

Product: FFP3 NR Mask
Plastic clips included
Stock Location EU
Individually Packed
CE certified FFP3 NR Masks

Color: White
Shelf Life: 3 years
Mask Dimensions: 15.5cm x 10.5cm

20 Mask per Box
50 Boxes per Carton
1,000 Masks per Carton

Carton Dimensions: 61x38x48cm
Carton Weight: 11kg

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

IRL: + 353 1 513 75 11
CH: + 41 71 58 80 248
www.OdemShop.com
info@OdemShop.com

Anhui Lekang Sanitary Materials Co., LTD

Qingcao Town Industrial Park, Tongcheng City, Anhui Province, China

---

DECLARATION

TO WHOM IT MAY CONCERN,

It is hereby certified and declared that the company:

“Better AG” located in General-Guisan-Str. 8, 6300 Zug, Switzerland

Is authorized to import, sell, distribute the "Xique" branded goods in Europe, Asia and Africa.

We hereby confirm the authenticity of the FFP2 & FFP3 masks sold by this distributor.

Signed for and on behalf of:

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Anhui Lekang Sanitary Materials CO.,Ltd.
Διεύθ.: Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China

ΔΗΛΩΣΗ ΣΥΜΜΟΡΦΩΣΗΣ ΕΕ

Ατομικός προστατευτικός εξοπλισμός:

Εταιρεία:	XIQUE
Όνομα:	Μάσκα φιλτραρίσματος μισού προσώπου
Μοντέλο:	LK-Z006
Πρότυπο:	EN 149:2001+A1:2009
Κλάση:	FFP2 NR
Χρώμα:	Άσπρο
Κατασκευαστής:-	Anhui lekang sanitary materials co., LTD
Διεύθυνση Κατασκευαστή:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, Κίνα

Η παρούσα δήλωση συμμόρφωσης εκδίδεται με αποκλειστική ευθύνη του κατασκευαστή:
Anhui Lekang sanitary materials Co., Ltd.
Το αντικείμενο της δήλωσης που περιγράφεται ανωτέρω είναι σύμφωνο με τη σχετική ενωσιακή νομοθεσία εναρμόνισης: Ατομικός προστατευτικός εξοπλισμός Κανονισμός (ΕΕ) 2016/425.

Έχει αποδειχθεί η εκπλήρωση των σχετικών απαιτήσεων υγείας και ασφάλειας που ορίζονται στο παράρτημα II.

Ο κοινοποιημένος οργανισμός:

LGAI TECHNOLOGICAL CENTER (APPLUS),
Campus UAB, Ronda de la Font del Carme s/n, E-08193 Bellaterra (Barcelona), Spain,
Αριθμός Κοινοποιημένου Οργανισμού: 0370

πραγματοποίησε την εξέταση τύπου ΕΕ (Ενότητα Β) και εξέδωσε το πιστοποιητικό εξέτασης τύπου ΕΕ με αριθμό κοινοποιημένου οργανισμού 0370-5639-PPE/B.

Το ΜΑΠ υπόκειται στη διαδικασία αξιολόγησης της συμμόρφωσης προς τον τύπο με βάση τον εσωτερικό έλεγχο της παραγωγής και τους εποπτευόμενους ελέγχους του προϊόντος σε τυχαία διαστήματα (ενότητα C2) που ορίζονται στον κανονισμό (ΕΕ) 2016/425, υπό την εποπτεία του κοινοποιημένου οργανισμού LGAI TECHNOLOGICAL CENTER (APPLUS), NB 0370.

Όνομα: XIN-LONG, LI
Θέση: Γενικός Διευθυντής
Τόπος: TONGCHENG
Ημερομηνία: 18/08/2021

FFP2 NR Fish Shape Mask

Anhui lekang sanitary materials co., LTD
Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China

EU DECLARATION OF CONFORMITY

Personal Protective Equipment:

Brand:	XIQUE
Name:	Filtering half mask
Model:	LK-L005
Standard:	EN 149:2001+A1:2009
Class:	FFP2 NR
Manufacturer:	Anhui lekang sanitary materials co., LTD
Manufacturer Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province,China

This declaration of conformity is issued under the sole responsibility of the manufacturer:
Anhui lekang sanitary materials co., LTD
The object of the declaration described above is in conformity with the relevant Union harmonization legislation: Personal Protective Equipment Regulation (EU) 2016/425.

The fulfillment of the relevant health and safety requirements set out in Annex II has been demonstrated.

The notify body:

LGAI TECHNOLOGICAL CENTER (APPLUS),
Campus UAB, Ronda de la Font del Carme s/n, E-08193 Bellaterra (Barcelona), Spain,
Notified Body Number:0370

performed the EU type-examination (Module B) and issued the EU type-examination certificate with notified body number 0370.

The PPE is subject to the conformity to type assessment procedure based on internal production control plus supervised product checks at random intervals (Module C2) set out in the Regulation (EU) 2016/425,under surveillance of the notified body LGAI TECHNOLOGICAL CENTER (APPLUS), NB 0370.

Signed for and on behalf of:
Name: Xinlong Li
Postion: Gerneral manager
Place: TONGCHENG
Date: 11/01/2022

FFP3 NR Mask

Anhui Lekang Sanitary Materials Co., LTD
Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China

EU DECLARATION OF CONFORMITY

Personal Protective Equipment:

Brand:	XIQUE
Name:	Filtering half mask
Model:	LK-T007
Standard:	EN 149:2001+A1:2009
Class:	FFP3 NR
Manufacturer:	Anhui lekang sanitary materials co., LTD
Manufacturer Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China

This declaration of conformity is issued under the sole responsibility of the manufacturer:
Anhui lekang sanitary materials co., LTD
The object of the declaration described above is in conformity with the relevant Union harmonization legislation: Personal Protective Equipment Regulation (EU) 2016/425.

The fulfillment of the relevant health and safety requirements set out in Annex II has been demonstrated.

The notify body:

LGAI TECHNOLOGICAL CENTER (APPLUS),
Campus UAB, Ronda de la Font del Carme s/n, E-08193 Bellaterra (Barcelona), Spain,
Notified Body Number:0370

performed the EU type-examination (Module B) and issued the EU type-examination certificate with notified body number 0370.

The PPE is subject to the conformity to type assessment procedure based on internal production control plus supervised product checks at random intervals (Module C2) set out in the Regulation (EU) 2016/425,under surveillance of the notified body LGAI TECHNOLOGICAL CENTER (APPLUS), NB 0370.

Signed for and on behalf of:

CE

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG

Top quality at manufacturer prices

FFP2 NR Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applusiabcratories.com

Applus

CE

Notified Body No. 0370

CERTIFICADO DE EXAMEN UE DE TIPO
EU-TYPE EXAMINATION CERTIFICATE

No. **0370- 5639-PPE/B**

ORGANISMO NOTIFICADO Nº *NOTIFIED BODY NUMBER*	**0370 - LGAI TECHNOLOGICAL CENTER** (APPLUS)
SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China.
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China.
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓN INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD *CONFORMITY ASSESSMENT PROCEDURE*	Módulo // *Module:* **B** **EXAMEN UE DE TIPO /** *EU TYPE EXAMINATION*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-2006 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
NORMAS ARMONIZADAS / *HARMONISED STANDARDS*	EN 149:2001 + A1:2009 Dispositivos de protección respiratoria. Medias máscaras filtrantes de protección contra partículas. Requisitos, ensayos, marcado. *EN 149:2001 + A1:2009 Respiratory protective devices. Filtering half masks to protect against particles. Requirements, testing, marking*
FECHA DE EMISIÓN / *ISSUE DATE*	**23/04/2021**
FECHA DE MODIFICACIÓN / *MODIFICATION DATE*	**12/01/2022**
VALIDEZ HASTA / *VALIDITY UNTIL*	**23/04/2026**

El presente certificado se mantendrá vigente durante 5 años siempre que el producto descrito no sea modificado y cumpla los requisitos esenciales de salud y seguridad establecidos en el Reglamento (UE) 2016/425. Para asegurar dicho cumplimiento, este certificado deberá ir acompañado de la documentación correspondiente a la Evaluación de Conformidad con el tipo según módulo C2, D (realizada por un Organismo Notificado, según frecuencia establecida).

This certificate will remain valid for 5 years as long as the indicated product is not modified and fulfills the essential requirements of health and safety established in (EU) Regulation 2016/425. To ensure such compliance, this certificate must be accompanied by the documentation corresponding to the Conformity Assessment to type according to C2, D(carried out by a Notified Body according, to the established frequency).

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC CERTIFICACIÓN

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusiaboratories.com/certified_products

Pagina 1 di 2

LGAI TECHNOLOGICAL CENTER, S.A. CIF: A-57157497

FFP2 NR Fish Shape Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applusiabcratories.com

Applus

CE

Notified Body No. 0370

CERTIFICADO DE EXAMEN UE DE TIPO
EU-TYPE EXAMINATION CERTIFICATE

No. **0370-5637-PPE/B**

ORGANISMO NOTIFICADO Nº *NOTIFIED BODY NUMBER*	**0370 - LGAI TECHNOLOGICAL CENTER** (APPLUS)
SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓN INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD *CONFORMITY ASSESSMENT PROCEDURE*	Módulo // *Module:* **B** **EXAMEN UE DE TIPO /** *EU TYPE EXAMINATION*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-L005 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
NORMAS ARMONIZADAS / *HARMONISED STANDARDS*	EN 149:2001 + A1:2009 Dispositivos de protección respiratoria. Medias máscaras filtrantes de protección contra partículas. Requisitos, ensayos, marcado. *EN 149:2001 + A1:2009 Respiratory protective devices. Filtering half masks to protect against particles. Requirements, testing, marking*
FECHA DE EMISIÓN / *ISSUE DATE*	**23/04/2021**
VALIDEZ HASTA / *VALIDITY UNTIL*	**23/04/2026**

El presente certificado se mantendrá vigente durante 5 años siempre que el producto descrito no sea modificado y cumpla los requisitos esenciales de salud y seguridad establecidos en el Reglamento (UE) 2016/425. Para asegurar dicho cumplimiento, este certificado deberá ir acompañado de la documentación correspondiente a la Evaluación de Conformidad con el tipo según módulo C2, D (realizada por un Organismo Notificado, según frecuencia establecida).

This certificate will remain valid for 5 years as long as the indicated product is not modified and fulfills the essential requirements of health and safety established in (EU) Regulation 2016/425. To ensure such compliance, this certificate must be accompanied by the documentation corresponding to the Conformity Assessment to type according to C2, D(carried out by a Notified Body according, to the established frequency).

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC CERTIFICACIÓN

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusiaboratories.com/certified_products

Pagina 1 di 2

LGAI TECHNOLOGICAL CENTER, S.A. CIF: A-57157497

FFP3 NR Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applusiabcratories.com

Applus

CE

Notified Body No. 0370

CERTIFICADO DE EXAMEN UE DE TIPO
EU-TYPE EXAMINATION CERTIFICATE

No. **0370-5638-PPE/B**

ORGANISMO NOTIFICADO Nº *NOTIFIED BODY NUMBER*	**0370 - LGAI TECHNOLOGICAL CENTER** (APPLUS)
SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓN INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD *CONFORMITY ASSESSMENT PROCEDURE*	Módulo // *Module:* **B** **EXAMEN UE DE TIPO /** *EU TYPE EXAMINATION*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-T007 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP3 NR
NORMAS ARMONIZADAS / *HARMONISED STANDARDS*	EN 149:2001 + A1:2009 Dispositivos de protección respiratoria. Medias máscaras filtrantes de protección contra partículas. Requisitos, ensayos, marcado. *EN 149:2001 + A1:2009 Respiratory protective devices. Filtering half masks to protect against particles. Requirements, testing, marking*
FECHA DE EMISIÓN / *ISSUE DATE*	**23/04/2021**
VALIDEZ HASTA / *VALIDITY UNTIL*	**23/04/2026**

El presente certificado se mantendrá vigente durante 5 años siempre que el producto descrito no sea modificado y cumpla los requisitos esenciales de salud y seguridad establecidos en el Reglamento (UE) 2016/425. Para asegurar dicho cumplimiento, este certificado deberá ir acompañado de la documentación correspondiente a la Evaluación de Conformidad con el tipo según módulo C2, D (realizada por un Organismo Notificado, según frecuencia establecida).

This certificate will remain valid for 5 years as long as the indicated product is not modified and fulfills the essential requirements of health and safety established in (EU) Regulation 2016/425. To ensure such compliance, this certificate must be accompanied by the documentation corresponding to the Conformity Assessment to type according to C2, D(carried out by a Notified Body according, to the established frequency).

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC CERTIFICACIÓN

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusiaboratories.com/certified_products

Pagina 1 di 2

LGAI TECHNOLOGICAL CENTER, S.A. CIF: A-57157497

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

IGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applauslaboratories.com

Applus

Technical Annex Ed. 2
04/01/2022

ANEXO TÉCNICO
TECHNICAL ANNEX

0370-5639-PPE/B

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

MARCA *BRAND*	XIQUE
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-2006 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
Color *Colour*	Blanco White
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO PLEGABLE VERTICAL, DE 5 CAPAS, CON LAZOS DE OREJA CON GANCHO Y CLIP NASAL INTERIOR. TAMAÑO: 155mm*105mm VALVELESS FILTERING HALF MASK, VERTICAL FOLDING TYPE, 5 LAYERS, WITH EARLOOPS WITH HOOK AND INTERIOR NOSE CLIP SIZE: 155mm*105mm
INFORME DE ENSAYO *TEST REPORT*	PTC21022204703C-EN01 issued by Precise Testing & Certification (Guangdong) Co.,Ltd.(PTC)

Color *Colour*	Negro Black
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO PLEGABLE VERTICAL, DE 5 CAPAS, CON LAZOS DE OREJA CON GANCHO Y CLIP NASAL INTERIOR. TAMAÑO: 155mm*105mm VALVELESS FILTERING HALF MASK, VERTICAL FOLDING TYPE, 5 LAYERS, WITH EARLOOPS WITH HOOK AND INTERIOR NOSE CLIP SIZE: 155mm*105mm
INFORME DE ENSAYO *TEST REPORT*	PTC21022204703C-EN01 issued by Precise Testing & Certification (Guangdong) Co.,Ltd.(PTC) & S21122802001E issued by Shenzhen NTEK Testing Technology Co., Ltd.

FFP2 NR Fish Shape Mask

IGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applauslaboratories.com

Applus

Technical Annex Ed. 1
23/04/2021

ANEXO TÉCNICO
TECHNICAL ANNEX

0370-5637-PPE/B

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

MARCA *BRAND*	XIQUE
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-L005 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO FISH PLEGABLE, DE 4 CAPAS, DE COLOR BLANCO, CON LAZOS DE OREJA CON GANCHO, CON ALMOHADILLA NASAL Y CLIP NASAL INTERIOR. TAMAÑO: Cuerpo: 205mm*80mm Paneles: 60mm&58mm VALVELESS FILTERING HALF MASK, FISH FOLDING TYPE, 4 LAYERS, WHITE COLOUR, WITH EARLOOPS WITH HOOK, WITH NOSE BRIDGE PAD AND INTERIOR NOSE CLIP SIZE: Main body: 205mm*80mm Panels: 60mm&58mm
INFORME DE ENSAYO *TEST REPORT*	PTC21022204701C-EN01 issued by Precise Testing & Certification (Guangdong) Co.,Ltd.(PTC)

FFP3 NR Mask

IGAI Technological Center, S.A. (APPLUS)
Campus UAB – Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
www.applauslaboratories.com

Applus

Technical Annex Ed. 1
23/04/2021

ANEXO TÉCNICO
TECHNICAL ANNEX

0370-5638-PPE/B

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

MARCA *BRAND*	XIQUE
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-T007 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP3 NR
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO PLEGABLE VERTICAL, DE 5 CAPAS, DE COLOR BLANCO, CON LAZOS DE OREJA CON GANCHO, CON ALMOHADILLA NASAL Y CLIP NASAL INTERIOR. TAMAÑO: 155mm*105mm VALVELESS FILTERING HALF MASK, VERTICAL FOLDING TYPE, 5 LAYERS, WHITE COLOUR, WITH EARLOOPS WITH HOOK, WITH NOSE BRIDGE PAD AND INTERIOR NOSE CLIP SIZE: 155mm*105mm
INFORME DE ENSAYO *TEST REPORT*	PTC21022204702C-EN01 issued by Precise Testing & Certification (Guangdong) Co.,Ltd.(PTC).

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG

Top quality at manufacturer prices

FFP2 NR Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB - Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-63207492
www.applusloratories.com

Applus

CERTIFICADO DE CONFORMIDAD CON EL TIPO
CONFORMITY TO TYPE CERTIFICATE
Organismo Notificado / *Notified Body Nr. 0370*

No. **0370-5852-PPE/C2**

SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓ INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD CON EL TIPO *CONFORMITY ASSESSMENT PROCEDURE TO TYPE*	Módulo // *Module:* **C2** **BASADA EN EL CONTROL INTERNO DE LA PRODUCCIÓN MÁS EL CONTROL SUPERVISADO DE LOS PRODUCTOS A INTERVALOS ALEATORIOS** *BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED CONTROL OF PRODUCTS AT ALEATORY INTERVALS*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-Z006 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
FECHA DE EMISIÓN / *ISSUE DATE*	**08/06/2022**
VALIDEZ HASTA / *VALIDITY UNTIL:*	**16/06/2023**

El presente certificado se mantendrá vigente durante 1 año siempre que no se modifiquen las condiciones establecidas en el Certificado de Examen UE de Tipo referenciado en el Anexo.
This certificate will remain in force for 1 year as long as the conditions established in the EU Type certificate referenced in the annex are not modified.

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusloratories.com/certified_products

El fabricante, después de completar la documentación técnica, el procedimiento de evaluación de la conformidad y la declaración UE de conformidad puede colocar el marcado CE bajo su responsabilidad / *The manufacturer, after the completion of the technical documentation, the conformity assessment procedures and the EU Declaration of conformity, may affix the CE Marking under his responsibility.*

CE

Page 1 / 2

FFP2 NR Fish Shape Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB - Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-63207492
www.applusloratories.com

Applus

CERTIFICADO DE CONFORMIDAD CON EL TIPO
CONFORMITY TO TYPE CERTIFICATE
Organismo Notificado / *Notified Body Nr. 0370*

No. **0370-5851-PPE/C2**

SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓ INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD CON EL TIPO *CONFORMITY ASSESSMENT PROCEDURE TO TYPE*	Módulo // *Module:* **C2** **BASADA EN EL CONTROL INTERNO DE LA PRODUCCIÓN MÁS EL CONTROL SUPERVISADO DE LOS PRODUCTOS A INTERVALOS ALEATORIOS** *BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED CONTROL OF PRODUCTS AT ALEATORY INTERVALS*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-L005 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
FECHA DE EMISIÓN / *ISSUE DATE*	**15/06/2022**
VALIDEZ HASTA / *VALIDITY UNTIL:*	**16/06/2023**

El presente certificado se mantendrá vigente durante 1 año siempre que no se modifiquen las condiciones establecidas en el Certificado de Examen UE de Tipo referenciado en el Anexo.
This certificate will remain in force for 1 year as long as the conditions established in the EU Type certificate referenced in the annex are not modified.

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusloratories.com/certified_products

El fabricante, después de completar la documentación técnica, el procedimiento de evaluación de la conformidad y la declaración UE de conformidad puede colocar el marcado CE bajo su responsabilidad / *The manufacturer, after the completion of the technical documentation, the conformity assessment procedures and the EU Declaration of conformity, may affix the CE Marking under his responsibility.*

CE

Page 1 / 2

FFP3 NR Mask

LGAI Technological Center, S.A. (APPLUS)
Campus UAB - Ronda de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-63207492
www.applusloratories.com

Applus

CERTIFICADO DE CONFORMIDAD CON EL TIPO
CONFORMITY TO TYPE CERTIFICATE
Organismo Notificado / *Notified Body Nr. 0370*

No. **0370-5850-PPE/C2**

SOLICITANTE / FABRICANTE *APPLICANT / MANUFACTURER*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
PLANTA DE PRODUCCIÓN *PRODUCTION SITE*	**Anhui lekang sanitary materials co., LTD** Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
REGLAMENTO DE APLICACIÓN PARA DAR LA CONFORMIDAD / *APPLICABLE REGULATION TO GIVE CONFORMITY:* **REGLAMENTO (UE) 2016/425 SOBRE LOS EQUIPOS DE PROTECCIÓ INDIVIDUAL** *REGULATION (EU) 2016/425 PERSONAL PROTECTIVE EQUIPMENT*	
PROCEDIMIENTO DE EVALUACIÓN DE LA CONFORMIDAD CON EL TIPO *CONFORMITY ASSESSMENT PROCEDURE TO TYPE*	Módulo // *Module:* **C2** **BASADA EN EL CONTROL INTERNO DE LA PRODUCCIÓN MÁS EL CONTROL SUPERVISADO DE LOS PRODUCTOS A INTERVALOS ALEATORIOS** *BASED ON INTERNAL PRODUCTION CONTROL PLUS SUPERVISED CONTROL OF PRODUCTS AT ALEATORY INTERVALS*
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER)*	Ref.: LK-T007 Filtering half mask
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP3 NR
FECHA DE EMISIÓN / *ISSUE DATE*	**08/06/2022**
VALIDEZ HASTA / *VALIDITY UNTIL:*	**16/06/2023**

El presente certificado se mantendrá vigente durante 1 año siempre que no se modifiquen las condiciones establecidas en el Certificado de Examen UE de Tipo referenciado en el Anexo.
This certificate will remain in force for 1 year as long as the conditions established in the EU Type certificate referenced in the annex are not modified.

Applus
LGAI Technological Center, S.A.
Xavier Ruiz Peña
Managing Director, Product Conformity B.U.

ENAC

Este documento carece de validez sin su anexo técnico, cuyo número coincide con el del certificado.
This document is not valid without its technical annex, whose number coincides with the number of certificate.

Puede comprobarse la validez de este certificado en nuestra página web / *You can check the validity of this certificate on our website:*
www.applusloratories.com/certified_products

El fabricante, después de completar la documentación técnica, el procedimiento de evaluación de la conformidad y la declaración UE de conformidad puede colocar el marcado CE bajo su responsabilidad / *The manufacturer, after the completion of the technical documentation, the conformity assessment procedures and the EU Declaration of conformity, may affix the CE Marking under his responsibility.*

CE

Page 1 / 2

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

LGAI Technological Center S.A. (APPLUS)
Campus UAB - Horta de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-83207492
www.applusiaboratories.com

Applus

Technical Annex Ed.2
19/01/2022

ANEXO TÉCNICO
TECHNICAL ANNEX

0370- 5852-PPE/C2

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

Nº CERTIFICADO DE EXAMEN UE DE TIPO *NR. EU TYPE EXAMINATION CERTIFICATE*	**0370-5639-PPE/B**
EMITIDO POR *ISSUED BY*	LGAI TECHNOLOGICAL CENTER S.A. **(Organismo notificado nº 0370 /** *Notified Body nr. 0370***).**
FECHA EMISIÓN *ISSUE DATE*	23/04/2021
VALIDEZ HASTA *VALIDITY UNTIL*	23/04/2026
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER*	Ref.: LK-Z006 Filtering half mask
MARCA *BRAND*	XIQUE
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
COLOR *COLOUR*	Blanco White Negro Black Azul Blue Gris claro Light Grey Borgoña Burgundy
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO PLEGABLE VERTICAL, DE 5 CAPAS, CON LAZOS DE OREJA CON GANCHO Y CLIP NASAL INTERIOR. TAMAÑO: 155mm*105mm VALVELESS FILTERING HALF MASK, VERTICAL FOLDING TYPE, 5 LAYERS, WITH EARLOOPS WITH HOOK AND INTERIOR NOSE CLIP SIZE: 155mm*105mm
INFORME DE ENSAYO *TEST REPORT*	S22052305803E issued by Shenzhen NTEK Testing Technology Co., Ltd.

FFP2 NR Fish Shape Mask

LGAI Technological Center S.A. (APPLUS)
Campus UAB - Horta de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-83207492
www.applusiaboratories.com

Applus

Technical Annex Ed.1
16/06/2021

ANEXO TÉCNICO
TECHNICAL ANNEX

0370- 5851-PPE/C2

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

Nº CERTIFICADO DE EXAMEN UE DE TIPO *NR. EU TYPE EXAMINATION CERTIFICATE*	**0370-5637-PPE/B**
EMITIDO POR *ISSUED BY*	LGAI TECHNOLOGICAL CENTER S.A. **(Organismo notificado nº 0370 /** *Notified Body nr. 0370***).**
FECHA EMISIÓN *ISSUE DATE*	23/04/2021
VALIDEZ HASTA *VALIDITY UNTIL*	23/04/2026
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER*	Ref.: LK-L005 Filtering half mask
MARCA *BRAND*	XIQUE
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI / *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP2 NR
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO FISH PLEGABLE, DE 4 CAPAS, DE COLOR BLANCO, CON LAZOS DE OREJA CON GANCHO, CON ALMOHADILLA NASAL Y CLIP NASAL INTERIOR. TAMAÑO: Cuerpo: 205mm*80mm Paneles: 60mm&58mm VALVELESS FILTERING HALF MASK, FISH FOLDING TYPE, 4 LAYERS, WHITE COLOUR, WITH EARLOOPS WITH HOOK, WITH NOSE BRIDGE PAD AND INTERIOR NOSE CLIP SIZE: Main body: 205mm*80mm Panels: 60mm&58mm
INFORME DE ENSAYO *TEST REPORT*	S22052305801E issued by Shenzhen NTEK Testing Technology Co., Ltd.

FFP3 NR Mask

LGAI Technological Center S.A. (APPLUS)
Campus UAB - Horta de la Font del Carme s/n
08193 Bellaterra (Barcelona)
T +34 93 567 20 00
CIF A-83207492
www.applusiaboratories.com

Applus

Technical Annex Ed.1
16/06/2021

ANEXO TÉCNICO
TECHNICAL ANNEX

0370- 5850-PPE/C2

I. MODELOS INCLUIDOS EN EL CERTIFICADO

REFERENCES INCLUDED IN THIS CERTIFICATE

Nº CERTIFICADO DE EXAMEN UE DE TIPO *NR. EU TYPE EXAMINATION CERTIFICATE*	**0370-5638-PPE/B**
EMITIDO POR *ISSUED BY*	LGAI TECHNOLOGICAL CENTER S.A. **(Organismo notificado nº 0370 /** *Notified Body nr. 0370***).**
FECHA EMISIÓN *ISSUE DATE*	23/04/2021
VALIDEZ HASTA *VALIDITY UNTIL*	23/04/2026
IDENTIFICACIÓN DEL EPI (NÚMERO DE TIPO) *IDENTIFICATION OF THE PPE (TYPE NUMBER*	Ref.: LK-T007 Filtering half mask
MARCA *BRAND*	XIQUE
NIVEL O NIVELES DE RENDIMIENTO O LA CLASE DE PROTECCIÓN DEL EPI *PERFORMANCE LEVEL OR PROTECTION CLASS OF THE PPE*	FFP3 NR
DESCRIPCIÓN *DESCRIPTION*	MEDIA MÁSCARA FILTRANTE, DE TIPO PLEGABLE VERTICAL, DE 5 CAPAS, DE COLOR BLANCO, CON LAZOS DE OREJA CON GANCHO, CON ALMOHADILLA NASAL Y CLIP NASAL INTERIOR. TAMAÑO: 155mm*105mm VALVELESS FILTERING HALF MASK, VERTICAL FOLDING TYPE, 5 LAYERS, WHITE COLOUR, WITH EARLOOPS WITH HOOK, WITH NOSE BRIDGE PAD AND INTERIOR NOSE CLIP SIZE: 155mm*105mm
INFORME DE ENSAYO *TEST REPORT*	S22052305802E issued by Shenzhen NTEK Testing Technology Co., Ltd.

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

中国认可
国际互认
检测
TESTING
CNAS L5772

Test Report

Test Requirement:
According to the requirement of the Module C2 (SPC CE-062_EN M5 PPE) of Applus+, the test item(s) of the sample is according to the standard EN149:2001+A1:2009.

Product Name:	Filtering half mask
Report No.:	PTC21022204706C-EN01
APPLUS +With ID number:	21/32302741
Client:	Anhui lekang sanitary materials co., LTD
Client Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Manufacturer:	Anhui lekang sanitary materials co., LTD
Manufacturer Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Contact:	Xinlong Li
Model(s):	LK-Z006
Classification:	FFP2 NR
Date of Tests:	2021.06.04~2021.06.10

Signed for and on Behalf of PTC

Prepare by: Checked by: Approved by:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

中国认可
国际互认
检测
TESTING
CNAS L5772

Test Report

EN 149:2001+A1:2009 protective devices. Filtering half masks to protect against particles. Requirements, testing, marking

Product:	Filtering half mask
Report No.:	PTC21022204701C-EN01
Client:	Anhui lekang sanitary materials co., LTD
Client Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Manufacturer:	Anhui lekang sanitary materials co., LTD
Manufacturer Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Contact:	Xinlong Li
Model(s):	LK-L005
Classification:	FFP2 NR
Date of Tests:	2021.02.23~2021.03.01

Signed for and on Behalf of PTC

Prepare by: Checked by: Approved by:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

中国认可
国际互认
检测
TESTING
CNAS L5772

Test Report

EN 149:2001+A1:2009 protective devices. Filtering half masks to protect against particles. Requirements, testing, marking

Product:	Filtering half mask
Report No.:	PTC21022204702C-EN01
Client:	Anhui lekang sanitary materials co., LTD
Client Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Manufacturer:	Anhui lekang sanitary materials co., LTD
Manufacturer Address:	Qingcaozhen Town Industrial Park, Tongcheng City, Anhui Province, China
Contact:	Xinlong Li
Model(s):	LK-T007
Classification:	FFP3 NR
Date of Tests:	2021.02.23~2021.03.01

Signed for and on Behalf of PTC

Prepare by: Checked by: Approved by:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 2 of 13

Summary of assessment

Clause	Assessment
7.3 Visual inspection	PASS
7.5 Material	PASS
7.9.1 Total inward leakage	PASS
7.9.2 Penetration of filter material	PASS
7.12 Carbon dioxide content of the inhalation air	PASS
7.16 Breathing resistance	PASS
7.18 Demountable parts	PASS

Remark:
PASS: comply with requirement of standard

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 2 of 14

Summary of assessment

Clause	Assessment
7.3 Visual inspection	NOT TESTED
7.4 Packaging	PASS
7.5 Material	PASS
7.6 Cleaning and disinfecting	N/A
7.7 Practical performance	PASS
7.8 Finish of parts	PASS
7.9.1 Total inward leakage	PASS
7.9.2 Penetration of filter material	PASS
7.10 Compatibility with skin	PASS
7.11 Flammability	PASS
7.12 Carbon dioxide content of the inhalation air	PASS
7.13 Head harness	PASS
7.14 Field of vision	PASS
7.15 Exhalation valve	N/A
7.16 Breathing resistance	PASS
7.17 Clogging	N/A
7.18 Demountable parts	PASS
9 Marking	NOT TESTED

Remark:
PASS: comply with requirement of standard
N/A: not application
NOT TESTED: the clause were not required

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 2 of 14

Summary of assessment

Clause	Assessment
7.3 Visual inspection	NOT TESTED
7.4 Packaging	PASS
7.5 Material	PASS
7.6 Cleaning and disinfecting	N/A
7.7 Practical performance	PASS
7.8 Finish of parts	PASS
7.9.1 Total inward leakage	PASS
7.9.2 Penetration of filter material	PASS
7.10 Compatibility with skin	PASS
7.11 Flammability	PASS
7.12 Carbon dioxide content of the inhalation air	PASS
7.13 Head harness	PASS
7.14 Field of vision	PASS
7.15 Exhalation valve	N/A
7.16 Breathing resistance	PASS
7.17 Clogging	N/A
7.18 Demountable parts	PASS
9 Marking	NOT TESTED

Remark:
PASS: comply with requirement of standard
N/A: not application
NOT TESTED: the clause were not required

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 3 of 13

Test Result:

Requirement	Test Result	Conclusion
7.3 Visual inspection The visual inspection shall also include the marking and the information supplied by the manufacturer.	Comply	**Pass**
7.5 Material Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used. Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer. After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps. When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	No mechanical failure after undergoing the conditioning described in 8.3.1, No collapse when conditioned in accordance with 8.3.1 and 8.3.2.	**Pass**
7.9.1 Total inward leakage For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than 25 % for FFP1, 11 % for FFP2, 5 % for FFP3 and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than 22 % for FFP1, 8 % for FFP2, 2 % for FFP3.	FFP2, Test results are shown in Annex A Table 7.9.1-A&B	**Pass**
7.9.2 Penetration of filter material The penetration of the filter of the particle filtering half mask shall meet the requirements of Table 1.	FFP2, Test results are shown in Annex A Table 7.9.2.	**Pass**

	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	≤ 20%	≤ 20%
FFP2	≤ 6%	≤ 6%
FFP3	≤ 1%	≤ 1%

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 3 of 14

Test Result:

Requirement	Test Result	Conclusion
7.3 Visual inspection The visual inspection shall also include the marking and the information supplied by the manufacturer.	Not tested	**Not tested**
7.4 Packaging Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	In accordance with the requirement.	**Pass**
7.5 Material Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used. Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer. After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps. When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	No mechanical failure after undergoing the conditioning described in 8.3.1, No collapse when conditioned in accordance with 8.3.1 and 8.3.2.	**Pass**
7.6 Cleaning and disinfecting If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer.	Single shift use only	**N/A**
7.7 Practical performance The particle filtering half mask shall undergo practical performance tests under realistic conditions	No imperfections	**Pass**
7.8 Finish of parts Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	No sharp edges or burrs.	**Pass**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 3 of 14

Test Result:

Requirement	Test Result	Conclusion
7.3 Visual inspection The visual inspection shall also include the marking and the information supplied by the manufacturer.	Not tested	**Not tested**
7.4 Packaging Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	In accordance with the requirement.	**Pass**
7.5 Material Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used. Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer. After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps. When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	No mechanical failure after undergoing the conditioning described in 8.3.1, No collapse when conditioned in accordance with 8.3.1 and 8.3.2.	**Pass**
7.6 Cleaning and disinfecting If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer.	Single shift use only	**N/A**
7.7 Practical performance The particle filtering half mask shall undergo practical performance tests under realistic conditions	No imperfections	**Pass**
7.8 Finish of parts Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	No sharp edges or burrs.	**Pass**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 4 of 13

7.12 Carbon dioxide content of the inhalation air
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0 % (by volume)

Test results are shown in Annex A Table 7.12. **Pass**

7.16 Breathing resistance

Classification	Maximum permitted resistance (mbar)		
	Inhalation		Exhalation
	30 l/min	95 l/min	160 l/min
FFP1	0.6	2.1	3.0
FFP2	0.7	2.4	3.0
FFP3	1.0	3.0	3.0

FFP2. Test results are shown in Annex A Table 7.16. **Pass**

7.18 Demountable parts
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand

Comply **Pass**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 4 of 14

7.9.1 Total inward leakage
For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than 25 % for FFP1, 11 % for FFP2, 5 % for FFP3

and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than 22 % for FFP1, 8 % for FFP2, 2 % for FFP3.

FFP2, Test results are shown in Annex A Table 7.9.1-A&B **Pass**

7.9.2 Penetration of filter material
The penetration of the filter of the particle filtering half mask shall meet the requirements of Table 1.

	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	≤ 20%	≤ 20%
FFP2	≤ 6%	≤ 6%
FFP3	≤ 1%	≤ 1%

FFP2 , Test results are shown in Annex A Table 7.9.2. **Pass**

7.10 Compatibility with skin
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.

No irritation or any other adverse effect to health. **Pass**

7.11 Flammability
When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5 s after removal from the flame.

Test results are shown in Annex A Table 7.11. **Pass**

7.12 Carbon dioxide content of the inhalation air
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume)

Test results are shown in Annex A Table 7.12. **Pass**

7.13 Head harness
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.

The head harness shall be adjustable or self-adjusting and shall be

Head harness can be donned and removed easily, adjustable or self-adjusting and **Pass**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 4 of 14

7.9.1 Total inward leakage
For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than 25 % for FFP1, 11 % for FFP2, 5 % for FFP3

and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than 22 % for FFP1, 8 % for FFP2, 2 % for FFP3.

FFP3, Test results are shown in Annex A Table 7.9.1-A&B **Pass**

7.9.2 Penetration of filter material
The penetration of the filter of the particle filtering half mask shall meet the requirements of Table 1.

	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	≤ 20%	≤ 20%
FFP2	≤ 6%	≤ 6%
FFP3	≤ 1%	≤ 1%

FFP3 , Test results are shown in Annex A Table 7.9.2. **Pass**

7.10 Compatibility with skin
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.

No irritation or any other adverse effect to health. **Pass**

7.11 Flammability
When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5 s after removal from the flame.

Test results are shown in Annex A Table 7.11. **Pass**

7.12 Carbon dioxide content of the inhalation air
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1,0 % (by volume)

Test results are shown in Annex A Table 7.12. **Pass**

7.13 Head harness
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.

The head harness shall be adjustable or self-adjusting and shall be

Head harness can be donned and removed easily, adjustable or self-adjusting and **Pass**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 5 of 13

Annex A: Summarization of Test Data

Table 7.9.1-A: Inward Leakage Test Data

Test specification: EN 149:2001+A1:2009 Clause 8.5

Subject	Sample No.	Condition	Walk (%)	Head Side/side (%)	Head up/down (%)	Talk (%)	Walk (%)	Mean (%)
Lv	1	A.R	3.5	3.5	4.4	4.1	4.2	3.9
Li	2	A.R	3.7	4.0	4.3	3.9	3.8	3.9
Zhong	3	A.R	3.2	3.6	3.2	3.5	3.4	3.4
Xu	4	A.R	3.8	4.2	4.2	4.3	3.6	4.0
Ma	5	A.R	3.5	3.2	3.3	3.6	3.3	3.4
Chen	6	T.C	4.0	4.2	4.2	4.4	4.4	4.2
Chen	7	T.C	3.9	4.1	4.1	4.5	4.0	4.1
Zhuo	8	T.C	4.2	4.6	4.4	4.7	4.2	4.4
Chen	9	T.C	3.7	3.8	3.9	4.0	4.5	4.0
Zhang	10	T.C	3.4	3.5	3.7	3.7	3.5	3.6

Table 7.9.1-B: Facial dimension

Subject	Face Length	Face Width	Face Depth	Mouth Width
Lv	113	139	104	53
Li	120	135	112	55
Zhong	108	135	106	56
Xu	120	150	120	70
Ma	130	170	130	80
Chen	110	160	90	40
Chen	115	145	110	50
Zhuo	103	146	100	50
Chen	110	145	95	40
Zhang	144	141	101	54

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 5 of 14

sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	have sufficiently robust to hold the particle filtering half mask firmly.	
7.14 Field of vision The field of vision is acceptable if determined so in practical performance tests.	Pass the practical performance tests.	**Pass**
7.15 Exhalation valve A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations. If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9. Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s. When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10 N applied for 10 s.	No exhalation valve	**N/A**

7.16 Breathing resistance

Classification	Maximum permitted resistance (mbar)		
	Inhalation		Exhalation
	30 l/min	95 l/min	160 l/min
FFP1	0.6	2.1	3.0
FFP2	0.7	2.4	3.0
FFP3	1.0	3.0	3.0

FFP2. Test results are shown in Annex A Table 7.16. **Pass**

7.17 Clogging **7.17.2 Breathing resistance** Valved particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1: 4 mbar, FFP2: 5 mbar, FFP3: 7 mbar at 95L/min continuous flow The exhalation resistance shall not exceed 3 mbar at 160 L/min	Single shift use only.	**N/A**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 5 of 14

sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	have sufficiently robust to hold the particle filtering half mask firmly.	
7.14 Field of vision The field of vision is acceptable if determined so in practical performance tests.	Pass the practical performance tests.	**Pass**
7.15 Exhalation valve A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations. If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9. Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s. When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10 N applied for 10 s.	No exhalation valve	**N/A**

7.16 Breathing resistance

Classification	Maximum permitted resistance (mbar)		
	Inhalation		Exhalation
	30 l/min	95 l/min	160 l/min
FFP1	0.6	2.1	3.0
FFP2	0.7	2.4	3.0
FFP3	1.0	3.0	3.0

FFP3. Test results are shown in Annex A Table 7.16. **Pass**

7.17 Clogging **7.17.2 Breathing resistance** Valved particle filtering half masks: After clogging the inhalation resistances shall not exceed: FFP1: 4 mbar, FFP2: 5 mbar, FFP3: 7 mbar at 95L/min continuous flow The exhalation resistance shall not exceed 3 mbar at 160 L/min	Single shift use only.	**N/A**

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 6 of 13

Table 7.9.2: Penetration of filter material

Test specification: EN 149:2001+A1:2009 Clause 8.11

Aerosol	Condition	Sample No.	Penetration (%)	Assessment
Sodium chloride test	As received	11	0.3	Pass
		12	0.3	
		13	0.2	
	Simulated wearing treatment	14	0.5	
		15	0.4	
		16	0.4	
	Mechanical strength + Temperature conditioned	17	0.7	
		18	0.6	
		19	0.7	
Paraffin oil test	As received	20	0.2	
		21	0.3	
		22	0.2	
	Simulated wearing treatment	23	0.4	
		24	0.6	
		25	0.4	
	Mechanical strength + Temperature conditioned	26	0.7	
		27	0.3	
		28	0.3	

Table 7.12: Carbon dioxide content of the inhalation air

Test specification: EN 149:2001+A1:2009 Clause 8.7

Condition	Sample No.	Result (%)		Assessment
As received	29	0.05	Mean value: 0.04	Pass
	30	0.04		
	31	0.04		

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 6 of 14

continuous flow

Valveless particle filtering half masks
After clogging the inhalation and exhalation resistances shall not exceed:
FFP1: 3 mbar, FFP2: 4 mbar, FFP3: 5 mbar at 95L/min continuous flow

7.17.3 Penetration of filter material

	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	≤ 20%	≤ 20%
FFP2	≤ 6%	≤ 6%
FFP3	≤ 1%	≤ 1%

7.18 Demountable parts
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand — Comply — **Pass**

9 Marking

9.1 Packaging
The following information shall be clearly and durably marked on the smallest commercially available packaging or legible through it if the packaging is transparent.

9.1.1 The name, trademark or other means of identification of the manufacturer or supplier.
9.1.2 Type-identifying marking.
9.1.3 Classification
The appropriate class (FFP1, FFP2 or FFP3) followed by a single space and then: "NR" if the particle filtering half mask is limited to single shift use only. Example: FFP3 NR, or "R" if the particle filtering half mask is re-usable.
Example: FFP2 R D. — Not tested — **Not tested**
9.1.4 The number and year of publication of this European Standard.
9.1.5 At least the year of end of shelf life. The end of shelf life may be informed by a pictogram as shown in Figure 12a, where yyyy/mm indicates the year and month.
9.1.6 The sentence 'see information supplied by the manufacturer', at least in the official language(s) of the country of destination, or by using

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 6 of 14

continuous flow

Valveless particle filtering half masks
After clogging the inhalation and exhalation resistances shall not exceed:
FFP1: 3 mbar, FFP2: 4 mbar, FFP3: 5 mbar at 95L/min continuous flow

7.17.3 Penetration of filter material

	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	≤ 20%	≤ 20%
FFP2	≤ 6%	≤ 6%
FFP3	≤ 1%	≤ 1%

7.18 Demountable parts
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand — Comply — **Pass**

9 Marking

9.1 Packaging
The following information shall be clearly and durably marked on the smallest commercially available packaging or legible through it if the packaging is transparent.

9.1.1 The name, trademark or other means of identification of the manufacturer or supplier.
9.1.2 Type-identifying marking.
9.1.3 Classification
The appropriate class (FFP1, FFP2 or FFP3) followed by a single space and then: "NR" if the particle filtering half mask is limited to single shift use only. Example: FFP3 NR, or "R" if the particle filtering half mask is re-usable.
Example: FFP2 R D. — Not tested — **Not tested**
9.1.4 The number and year of publication of this European Standard.
9.1.5 At least the year of end of shelf life. The end of shelf life may be informed by a pictogram as shown in Figure 12a, where yyyy/mm indicates the year and month.
9.1.6 The sentence 'see information supplied by the manufacturer', at least in the official language(s) of the country of destination, or by using

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 7 of 13

Table 7.16: Breathing resistance (mbar)

Test specification: EN 149:2001+A1:2009 Clause 8.9

As received	Flow Rate		32					33					34					
	Inhalation	30 l/min	0.40					0.39					0.40					
		95 l/min	1.26					1.32					1.29					
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E	
			1.92	1.92	1.87	1.88	1.89	1.84	1.83	1.84	1.84	1.81	1.93	1.88	1.88	1.87	1.87	
Simulated wearing treatment	Flow Rate		35					36					37					
	Inhalation	30 l/min	0.36					0.39					0.38					
		95 l/min	1.32					1.28					1.26					
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E	
			1.86	1.81	1.77	1.80	1.83	1.89	1.85	1.87	1.89	1.87	1.87	1.85	1.85	1.82	1.82	
Temperature conditioned	Flow Rate		38					39					40					
	Inhalation	30 l/min	0.33					0.35					0.36					
		95 l/min	1.14					1.14					1.14					
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E	
			1.67	1.63	1.63	1.66	1.62	1.65	1.61	1.63	1.59	1.62	1.66	1.66	1.66	1.65	1.64	
Assessment	Pass																	

A: Facing directly ahead B: Facing vertically upwards C: Facing vertically downwards

D: Lying on the left side E: Lying on the right side

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 7 of 14

the pictogram as shown in Figure 12b.
9.1.7 The manufacturer's recommended conditions of storage (at least the temperature and humidity) or equivalent pictogram, as shown in Figures 12c and 12d.
9.1.8 The packaging of those particle filtering half masks passing the dolomite clogging test shall be additionally marked with the letter "D". This letter shall follow the classification marking preceded by a single space.

9.2 Particle filtering half mask
Particle filtering half masks complying with this European Standard shall be clearly and durably marked with the following:
9.2.1 The name, trademark or other means of identification of the manufacturer or supplier.
9.2.2 Type-identifying marking.
9.2.3 The number and year of publication of this European Standard.
9.2.4 Classification
The appropriate class (FFP1, FFP2 or FFP3) followed by a single space and then: "NR" if the particle filtering half mask is limited to single shift use only. Example: FFP3 NR, or "R" if the particle filtering half mask is re-usable. Example: FFP2 R D.
9.2.5 If appropriate the letter D (dolomite) in accordance with clogging performance. This letter shall follow the classification marking preceded by a single space.
9.2.6 Sub-assemblies and components with considerable bearing on safety shall be marked so that they can be identified.

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 7 of 14

the pictogram as shown in Figure 12b.
9.1.7 The manufacturer's recommended conditions of storage (at least the temperature and humidity) or equivalent pictogram, as shown in Figures 12c and 12d.
9.1.8 The packaging of those particle filtering half masks passing the dolomite clogging test shall be additionally marked with the letter "D". This letter shall follow the classification marking preceded by a single space.

9.2 Particle filtering half mask
Particle filtering half masks complying with this European Standard shall be clearly and durably marked with the following:
9.2.1 The name, trademark or other means of identification of the manufacturer or supplier.
9.2.2 Type-identifying marking.
9.2.3 The number and year of publication of this European Standard.
9.2.4 Classification
The appropriate class (FFP1, FFP2 or FFP3) followed by a single space and then: "NR" if the particle filtering half mask is limited to single shift use only. Example: FFP3 NR, or "R" if the particle filtering half mask is re-usable. Example: FFP2 R D.
9.2.5 If appropriate the letter D (dolomite) in accordance with clogging performance. This letter shall follow the classification marking preceded by a single space.
9.2.6 Sub-assemblies and components with considerable bearing on safety shall be marked so that they can be identified.

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 8 of 13

Test	Uncertainty
Total inward leakage	3.8%
Penetration of filter material(NaCl)	3.5%
Penetration of filter material(Paraffin oil)	4.2%
Carbon dioxide content of the inhalation air	4.5%
Breathing resistance(30L/min)	5.2%
Breathing resistance(95L/min)	5.4%
Breathing resistance(160)L/min)	6.0%

Photo(s) of Sample:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission.

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 8 of 14

Annex A: Summarization of Test Data

Table 7.9.1-A: Inward Leakage Test Data

Test specification：EN 149:2001+A1:2009 Clause 8.5

Subject	Sample No.	Condition	Walk (%)	Head Side/side (%)	Head up/down (%)	Talk (%)	Walk (%)	Mean (%)
Lv	1	A.R	3.5	3.3	3.5	4.9	4.1	3.9
Li	2	A.R	3.4	4.5	4.7	4.7	4.5	4.4
Zhong	3	A.R	4.3	4.5	6.1	6.0	4.5	5.1
Xu	4	A.R	3.1	3.1	3.4	3.5	3.4	3.3
Ma	5	A.R	4.4	4.9	4.8	5.0	5.0	4.8
Chen	6	T.C	3.8	3.7	4.5	4.6	4.3	4.2
Chen	7	T.C	3.8	3.5	3.4	3.5	3.4	3.5
Zhuo	8	T.C	4.3	4.6	4.8	4.5	5.0	4.6
Chen	9	T.C	3.5	3.2	4.4	3.5	3.1	3.5
Zhang	10	T.C	4.3	4.5	4.9	5.0	5.1	4.8

Table 7.9.1-B: Facial dimension

Subject	Face Length	Face Width	Face Depth	Mouth Width
Lv	113	139	104	53
Li	120	135	112	55
Zhong	108	135	106	56
Xu	120	150	120	70
Ma	130	170	130	80
Chen	110	160	90	40
Chen	115	145	110	50
Zhuo	103	146	100	50
Chen	110	145	95	40
Zhang	144	141	101	54

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 8 of 14

Annex A: Summarization of Test Data

Table 7.9.1-A: Inward Leakage Test Data

Test specification：EN 149:2001+A1:2009 Clause 8.5

Subject	Sample No.	Condition	Walk (%)	Head Side/side (%)	Head up/down (%)	Talk (%)	Walk (%)	Mean (%)
Lv	1	A.R	2.1	2.2	2.3	2.3	2.3	2.2
Li	2	A.R	1.9	1.8	1.7	1.8	1.7	1.8
Zhong	3	A.R	1.6	1.7	1.6	1.6	1.7	1.6
Xu	4	A.R	1.6	1.9	1.9	1.8	1.8	1.8
Ma	5	A.R	1.7	1.6	1.5	1.5	1.6	1.6
Chen	6	T.C	1.8	1.8	1.9	1.7	1.4	1.7
Chen	7	T.C	1.6	1.7	1.6	1.8	1.8	1.7
Zhuo	8	T.C	1.4	1.6	1.7	1.9	1.6	1.6
Chen	9	T.C	1.7	1.9	1.5	1.9	1.6	1.7
Zhang	10	T.C	1.6	1.7	1.9	1.8	1.5	1.7

Table 7.9.1-B: Facial dimension

Subject	Face Length	Face Width	Face Depth	Mouth Width
Lv	113	139	104	53
Li	120	135	112	55
Zhong	108	135	106	56
Xu	120	150	120	70
Ma	130	170	130	80
Chen	110	160	90	40
Chen	115	145	110	50
Zhuo	103	146	100	50
Chen	110	145	95	40
Zhang	144	141	101	54

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted. The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 9 of 13



This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 9 of 14

Table 7.9.2: Penetration of filter material

Test specification：EN 149:2001+A1:2009 Clause 8.11

Aerosol	Condition	Sample No.	Penetration (%)	Assessment
Sodium chloride test	As received	11	0.1	Pass
		12	0.2	
		13	0.5	
	Simulated wearing treatment	14	1.0	
		15	0.6	
		16	0.2	
	Mechanical strength + Temperature conditioned	17	0.1	
		18	0.4	
		19	0.2	
Paraffin oil test	As received	20	0.1	
		21	0.6	
		22	0.6	
	Simulated wearing treatment	23	0.1	
		24	0.4	
		25	0.4	
	Mechanical strength + Temperature conditioned	26	0.1	
		27	0.1	
		28	0.1	

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 9 of 14

Table 7.9.2: Penetration of filter material

Test specification：EN 149:2001+A1:2009 Clause 8.11

Aerosol	Condition	Sample No.	Penetration (%)	Assessment
Sodium chloride test	As received	11	0.2	Pass
		12	0.2	
		13	0.3	
	Simulated wearing treatment	14	0.1	
		15	0.2	
		16	0.1	
	Mechanical strength + Temperature conditioned	17	0.1	
		18	0.3	
		19	0.2	
Paraffin oil test	As received	20	0.1	
		21	0.1	
		22	0.1	
	Simulated wearing treatment	23	0.1	
		24	0.1	
		25	0.1	
	Mechanical strength + Temperature conditioned	26	0.1	
		27	0.1	
		28	0.1	

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 10 of 13

PTC21022204706C

PTC21022204706C

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 10 of 14

Table 7.11: Flammability

Test specification：EN 149:2001+A1:2009 Clause 8.6

Condition	Sample No.	Result	Assessment
As received	29	No burn	Pass
	30	No burn	
Temperature conditioned	31	No burn	
	32	No burn	

Table 7.12: Carbon dioxide content of the inhalation air

Test specification：EN 149:2001+A1:2009 Clause 8.7

Condition	Sample No.	Result (%)		Assessment
As received	33	0.04	Mean value: 0.04	Pass
	34	0.04		
	35	0.04		

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 10 of 14

Table 7.11: Flammability

Test specification：EN 149:2001+A1:2009 Clause 8.6

Condition	Sample No.	Result	Assessment
As received	29	No burn	Pass
	30	No burn	
Temperature conditioned	31	No burn	
	32	No burn	

Table 7.12: Carbon dioxide content of the inhalation air

Test specification：EN 149:2001+A1:2009 Clause 8.7

Condition	Sample No.	Result (%)		Assessment
As received	33	0.04	Mean value: 0.04	Pass
	34	0.04		
	35	0.04		

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG

Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10,2021 Page 11 of 13

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 11 of 14

Table 7.16: Breathing resistance (mbar)

Test specification：EN 149:2001+A1:2009 Clause 8.9

As received	Flow Rate		36					37					38						
	Inhalation	30 l/min	0.31					0.31					0.34						
		95 l/min	1.27					1.29					1.29						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			2.13	2.12	2.13	2.13	2.13	2.10	2.12	2.11	2.12	2.13	2.10	2.12	2.09	2.12	2.10		
Simulated wearing treatment	Flow Rate		39					40					41						
	Inhalation	30 l/min	0.32					0.33					0.37						
		95 l/min	1.28					1.33					1.35						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			2.17	2.15	2.16	2.15	2.16	2.19	2.17	2.13	2.15	2.16	2.10	2.19	2.13	2.10	2.14		
Temperature conditioned	Flow Rate		42					43					44						
	Inhalation	30 l/min	0.33					0.31					0.33						
		95 l/min	1.12					1.14					1.13						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			1.74	1.68	1.71	1.68	1.66	1.70	1.68	1.68	1.70	1.70	1.69	1.68	1.70	1.68	1.67		
Assessment	Pass																		

A: Facing directly ahead B: Facing vertically upwards C: Facing vertically downwards

D: Lying on the left side E: Lying on the right side

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 11 of 14

Table 7.16: Breathing resistance (mbar)

Test specification：EN 149:2001+A1:2009 Clause 8.9

As received	Flow Rate		36					37					38						
	Inhalation	30 l/min	0.38					0.38					0.38						
		95 l/min	1.23					1.22					1.21						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			1.90	1.89	1.90	1.91	1.91	1.88	1.85	1.92	1.88	1.93	1.91	1.89	1.90	1.89	1.85		
Simulated wearing treatment	Flow Rate		39					40					41						
	Inhalation	30 l/min	0.34					0.36					0.32						
		95 l/min	1.30					1.31					1.28						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			2.12	2.09	2.09	2.11	2.11	2.04	2.08	2.07	2.05	2.03	2.11	2.08	2.07	2.07	2.05		
Temperature conditioned	Flow Rate		42					43					44						
	Inhalation	30 l/min	0.32					0.32					0.32						
		95 l/min	1.16					1.18					1.18						
	Exhalation	160 l/min	A	B	C	D	E	A	B	C	D	E	A	B	C	D	E		
			1.82	1.78	1.76	1.81	1.79	1.79	1.80	1.76	1.79	1.77	1.80	1.78	1.74	1.78	1.81		
Assessment	Pass																		

A: Facing directly ahead B: Facing vertically upwards C: Facing vertically downwards

D: Lying on the left side E: Lying on the right side

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10.2021 Page 12 of 13

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 12 of 14

Test	Uncertainty
Total inward leakage	3.8%
Penetration of filter material(NaCl)	3.5%
Penetration of filter material(Paraffin oil)	4.2%
Carbon dioxide content of the inhalation air	4.5%
Breathing resistance(30L/min)	5.2%
Breathing resistance(95L/min)	5.4%
Breathing resistance(160)L/min)	6.0%

Photo(s) of Sample:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 12 of 14

Test	Uncertainty
Total inward leakage	3.8%
Penetration of filter material(NaCl)	3.5%
Penetration of filter material(Paraffin oil)	4.2%
Carbon dioxide content of the inhalation air	4.5%
Breathing resistance(30L/min)	5.2%
Breathing resistance(95L/min)	5.4%
Breathing resistance(160)L/min)	6.0%

Photo(s) of Sample:

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

BETTER AG
Top quality at manufacturer prices

FFP2 NR Mask

Test Report

Report No.:PTC21022204706C-EN01 **Issue Date:** Jun.10.2021 Page 13 of 13



End of Report

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 13 of 14

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 13 of 14



This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

BETTER AG
General-Guisan-Str.8
6300 Zug, Switzerland

FFP2 NR Fish Shape Mask

Test Report

Report No.:PTC21022204701C-EN01 **Issue Date:** Mar.01, 2021 Page 14 of 14



End of Report

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com

FFP3 NR Mask

Test Report

Report No.:PTC21022204702C-EN01 **Issue Date:** Mar.01, 2021 Page 14 of 14

End of Report

This report is issued according to the general clauses in www.ptc-testing.com by PTC. Responsibility, guarantee and law restriction are defined in the general service clauses. The report is only responsible for the submitted sample(s) except as otherwise noted.
The report could not be copied without permission

Precise Testing & Certification (Guangdong) Co., Ltd. (PTC)
Building 1, No. 6, Tongxin Road, Dongcheng Street, Dongguan, Guangdong, China.
Tel: 86-769-38808222 Fax: 86-769-38826111 http:// www.ptc-testing.com